

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 9090			
SULFURIC ACID 1.0 NORMAL NIST TRACEABLE			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 04/27/23			
Tester JENC			
Recorded 04/27/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 04/27/23			
Tester JENC			
Recorded 04/27/23			
Actual Normality	1.0000	0.9995 - 1.0005	1.0002
	EQ/L		
Sample ID			
Tested 04/27/23			
Tester JENC			
Recorded 04/27/23			
QCSP5: STANDARDIZATION OF	YES		YES
AQUEOUS ACIDS	METHOD		
Sample ID			
Tested 04/27/23			
Tester JENC			
Recorded 04/27/23			
N.I.S.T. Traceable to SRM 723 THAM	YES		YES
	YES		
Sample ID			
Tested 04/27/23			
Tester JENC			
Recorded 04/27/23			
Actual Density	1.02999	1.02977 - 1.03021	1.03001
	GM/ML		
Sample ID			
Tested 04/27/23			
Tester JENC			
Recorded 04/27/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 04/27/23			
Tester JENC			
Recorded 04/27/23			
INSTRUMENTS USED DURING	INSTRUMENT		T-223,GC,T-211,M55,BID02253
PREPARATION			,PIDM91
Sample ID			
Tested 04/27/23			
Tester JENC			
Recorded 04/27/23			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 04/27/23 at 4:19 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

EXPIRATION DATE

MM/DD/YY
MM/DD/YYYY

04/30/25

Sample ID

Tested 04/27/23

Tester JENC

Recorded 04/27/23

SHELF LIFE OF PRODUCT

SHELF LIFE
TIME

2 YEARS

Sample ID

Tested 04/27/23

Tester JENC

Recorded 04/27/23

Lot # 3041980

Made 04/27/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager