

Page: 1

Date: 04/25/23 at 1:39 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	CY196F			
		STANDARD 0.2 PPM		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES	-	YES
FILE		YES		
Sample ID				
Tested	04/25/23			
Tester	AROS			
Recorded	04/25/23			
Calculated ppm Cyanide (CN)		0.200	0.190 - 0.210	0.199
		PPM W/V		
Sample ID				
Tested	04/25/23			
Tester	AROS			
Recorded	04/25/23			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	04/25/23			
Tester	AROS			
Recorded	04/25/23			
INSTRUMENTS USED D		INSTRUMENT		P92,P389,GC,VF3-J3
PREPARATION	Johna	INGTHOMENT		1 32,1 333,433,41 3 33
Sample ID				
Tested	04/25/23			
	04/25/23 AROS			
Tester				
Recorded	04/25/23			05/00/00
EXPIRATION DATE		MM/DD/YYYY		05/30/23
		MM/DD/YYYY		
Sample ID				
Tested	04/25/23			
Tester	AROS			
Recorded	04/25/23			
SHELF LIFE OF PRODU	ICT	1 MONTH TIME		1 MONTH
Comple ID		I IIVIL		
Sample ID Tested	04/05/00			
	04/25/23			
Tester	AROS			
Recorded	04/25/23			
Lot # 3042569				
Made 04/25	5/23			



Page: 2

Date: 04/25/23 at 1:39 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager