

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: 8818 SODIUM THIOSULFATE 0.5 NORMAL | | | |
|--|---|-------------------|-----------------------------|
| Test | Target/UOM | Range | Result |
| PREPARED TO FORMULATION ON | YES | | YES |
| FILE | YES | | |
| Sample ID | | | |
| Tested 04/27/23 | | | |
| Tester MHUS | | | |
| Recorded 04/27/23 | | | |
| Actual Normality | 0.5000 EQ/L | 0.4995 - 0.5005 | 0.4997 |
| Sample ID | | | |
| Tested 04/27/23 | | | |
| Tester MHUS | | | |
| Recorded 04/27/23 | | | |
| Actual Density | 1.06068 GM/ML | 1.05955 - 1.06180 | 1.06101 |
| Sample ID | | | |
| Tested 04/27/23 | | | |
| Tester MHUS | | | |
| Recorded 04/27/23 | | | |
| QCSP10:STANDARDIZATION OF | YES | | YES |
| SODIUM THIOSUL | QCSP10 | | |
| Sample ID | | | |
| Tested 04/27/23 | | | |
| Tester MHUS | | | |
| Recorded 04/27/23 | | | |
| N.I.S.T. Traceable to SRM 136 | YES | | YES |
| | YES | | |
| Sample ID | | | |
| Tested 04/27/23 | | | |
| Tester MHUS | | | |
| Recorded 04/27/23 | | | |
| APPEARANCE AND COLOR | CLEAR WATER WHITE LIQUID APPEARANCE | | PASS |
| Sample ID | | | |
| Tested 04/27/23 | | | |
| Tester MHUS | | | |
| Recorded 04/27/23 | | | |
| INSTRUMENTS USED DURING | INSTRUMENT | | M103,VF12-96,M55,BID3902,PI |
| PREPARATION | | | DM91 |
| Sample ID | | | |
| Tested 04/27/23 | | | |
| Tester MHUS | | | |
| Recorded 04/27/23 | | | |
| SHELF LIFE OF PRODUCT | 6 MONTHS TIME | | 6 MONTHS |
| Sample ID | | | |
| Tested 04/27/23 | | | |
| Tester MHUS | | | |
| Recorded 04/27/23 | | | |



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 04/27/23 at 4:17 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

EXPIRATION DATE

MM/DD/YY
MM/DD/YYYY

10/30/23

Sample ID

Tested 04/27/23

Tester MHUS

Recorded 04/27/23

Lot # 3041924

Made 04/27/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager