



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 1

Date: 04/24/23 at 1:47 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: LY386F SULFUR STANDARD 10.0 PPM W/W IN ISOOCTANE | | | |
|--|----------------------------|--------------|------------------|
| Test | Target/UOM | Range | Result |
| PREPARED TO FORMULATION ON | YES | | YES |
| FILE | YES | | |
| Sample ID | | | |
| Tested 04/24/23 | | | |
| Tester MMAT | | | |
| Recorded 04/24/23 | | | |
| Calculated ppm Sulfur | 10.00 PPM | 9.80 - 10.20 | 10.00 |
| Sample ID | | | |
| Tested 04/24/23 | | | |
| Tester MMAT | | | |
| Recorded 04/24/23 | | | |
| APPEARANCE AND COLOR | CLEAR LIQUID APPEARANCE | | PASS |
| Sample ID | | | |
| Tested 04/24/23 | | | |
| Tester MMAT | | | |
| Recorded 04/24/23 | | | |
| INSTRUMENTS USED DURING PREPARATION | INSTRUMENT | | VF100-K2,VF1-Z20 |
| Sample ID | | | |
| Tested 04/24/23 | | | |
| Tester MMAT | | | |
| Recorded 04/24/23 | | | |
| Lot # | 3042464 | | |
| Made | 04/24/23 | | |

Quality Manager