

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: THE510F			
Sodium Hydroxide 0.1562 Normal NIST Traceable Soln			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 04/27/23			
Tester JENC			
Recorded 04/27/23			
NORMALITY TO 4 DEC PLACES	0.1562	0.1559 - 0.1565	0.1562
	EQUIV/LITER		
Sample ID			
Tested 04/27/23			
Tester JENC			
Recorded 04/27/23			
N.I.S.T. Traceable to SRM 84	YES		YES
	YES		
Sample ID			
Tested 04/27/23			
Tester JENC			
Recorded 04/27/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 04/27/23			
Tester JENC			
Recorded 04/27/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		04/27/23
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 04/27/23			
Tester JENC			
Recorded 04/27/23			
AMBIENT TEMPERATURE DURING	25.0	15.0 - 35.0	21.0
SAMPLING DEGC	DEG C		
Sample ID			
Tested 04/27/23			
Tester JENC			
Recorded 04/27/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		04/27/23
	MM/DD/YYYY		
Sample ID			
Tested 04/27/23			
Tester JENC			
Recorded 04/27/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 25.0	21.0
DEG C	DEG C		
Sample ID			
Tested 04/27/23			
Tester JENC			
Recorded 04/27/23			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 04/27/23 at 8:16 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

APPEARANCE AND COLOR

CLEAR LIQUID
APPEARANCE

PASS

Sample ID
Tested 04/27/23
Tester JENC
Recorded 04/27/23

**INSTRUMENTS USED DURING
PREPARATION**

INSTRUMENT

M112,T-52A,BID-02253,PID-M9
1

Sample ID
Tested 04/27/23
Tester JENC
Recorded 04/27/23

EXPIRATION DATE

MM/DD/YYYY
MM/DD/YYYY

10/30/24

Sample ID
Tested 04/27/23
Tester JENC
Recorded 04/27/23

Lot # 3042112

Made 04/27/23

Quality Manager