

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8787 SODIUM THIOSULFATE 0.005 NORMAL SOLUTION			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE Sample ID Tested 04/26/23 Tester MHUS Recorded 04/26/23	YES YES		YES
NORMALITY TO 4 DEC PLACES Sample ID Tested 04/26/23 Tester MHUS Recorded 04/26/23	0.0050 EQUIV/LITER	0.0048 - 0.0052	0.0049
Actual Density Sample ID Tested 04/26/23 Tester MHUS Recorded 04/26/23	0.99885 GM/ML	0.99767 - 1.00004	0.99892
APPEARANCE AND COLOR Sample ID Tested 04/26/23 Tester MHUS Recorded 04/26/23	CLEAR LIQUID APPEARANCE		PASS
INSTRUMENTS USED DURING PREPARATION Sample ID Tested 04/26/23 Tester MHUS Recorded 04/26/23	INSTRUMENT		GC,VF3-J17,M55,BURET:04099
EXPIRATION DATE Sample ID Tested 04/26/23 Tester MHUS Recorded 04/26/23	MM/DD/YY MM/DD/YYYY		08/30/23
SHELF LIFE OF PRODUCT Sample ID Tested 04/26/23 Tester MHUS Recorded 04/26/23	SHELF LIFE TIME		4 MONTHS
Lot # 3042482 Made 04/27/23			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 04/27/23 at 11:27 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager