Page: 1

Date: 05/05/23 at 4:11 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	: VWR092F			
		e Standard 0.5 ppm		
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID				
Tested	05/05/23			
Tester	MMAT			
Recorded	05/05/23			
Actual ppm Fluoride		0.50	0.48 - 0.52	0.50
		PPM		
Sample ID				
Tested	05/05/23			
Tester	MMAT			
Recorded	05/05/23			
APPEARANCE AND CO	LOR	CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	05/05/23			
Tester	MMAT			
Recorded	05/05/23			
INSTRUMENTS USED I		INSTRUMENT		VF2-A6
PREPARATION				
Sample ID				
Tested	05/05/23			
Tester	MMAT			
Recorded	05/05/23			
EXPIRATION DATE	00/00/20	MM/DD/YYYY		05/30/25
		MM/DD/YYYY		33,33,23
Sample ID		, 22,		
Tested	05/05/23			
Tester	MMAT			
Recorded	05/05/23			
Lot # 3050549				
Made 05/05	5/23			
iviaue 55755	-,			

Quality Manager