

Page: 1

Date: 04/25/23 at 4:30 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	5625			
Ü		IC NITRATE 0.01 NORMAL A	AQUEOUS	
Test		Target/UOM	Range	Result
Actual Normality		0.0100	0.0095 - 0.0105	0.0101
		EQ/L		
Sample ID				
Tested	04/25/23			
Tester	CCOA			
Recorded	04/25/23			
QCSP14:MERCURIC NIT	TRATE	YES		YES
STANDARDIZATION		QCSP14		
Sample ID				
Tested	04/25/23			
Tester	CCOA			
Recorded	04/25/23			
PREPARED TO FORMU	LATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	04/25/23			
Tester	CCOA			
Recorded	04/25/23			
APPEARANCE AND CO		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID		7.1. 7. 27.11.11.102		
Tested	04/25/23			
Tester	CCOA			
Recorded	04/25/23			
INSTRUMENTS USED D		INSTRUMENT		GC,T-226,M55,BID00488,PIDM
PREPARATION		INSTITUMENT		91
				91
Sample ID	0.1/07/00			
Tested	04/25/23			
Tester	CCOA			
Recorded	04/25/23	0.00000	0.00000 4.00004	0.00007
Actual Density		0.99982	0.99960 - 1.00004	0.99967
		GM/ML		
Sample ID				
Tested	04/25/23			
Tester	CCOA			
Recorded	04/25/23			
EXPIRATION DATE		MM/DD/YY		04/30/24
		MM/DD/YYYY		
Sample ID				
Tested	04/25/23			
Tester	CCOA			
Recorded	04/25/23			
Lot # 30116	67			
Made 04/25				
IVIAGE 0 17 20	, = -			



Page: 2

Date: 04/25/23 at 4:30 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager