

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Package	ed Product:	6490			
0		NITROGEN	STANDARD 1000 PPM (1	ML=1MG N), CERTIFIED	
Test			Target/UOM	Range	Result
PREPARE	ED TO FORMULA	TION ON	YES	V	YES
FILE			YES		
	Sample ID				
	Tested 04	4/24/23			
	Tester M	IMAT			
	Recorded 04	4/24/23			
PPM NITROGEN NO DECIAML		AML	1000	980 - 1020	983
PLACES			PPM		
	Sample ID				
	Tested 04	4/24/23			
	Tester M	IMAT			
	Recorded 04	4/24/23			
APPEARANCE AND COLOR		R	CLEAR WATER WHITE		PASS
			LIQUID		
			APPEARANCE		
	Sample ID				
	Tested 04	4/24/23			
	Tester M	IMAT			
	Recorded 04	4/24/23			
INSTRUMENTS USED DURING		RING	INSTRUMENT		M115,GP,VF3-J17,M107
PREPARATION					
	Sample ID				
	•	4/24/23			
	Tester N	IMAT			
	Recorded 04	4/24/23			
EXPIRATION DATE			MM/DD/YY		07/30/23
			MM/DD/YYYY		
	Sample ID				
		4/24/23			
	Tester N	IMAT			
	Recorded 04	4/24/23			
Lo	ot # 3042469	9			
	ade 04/24/2	23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1