



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 9451			
TISAB SOLUTION APHA-EPA FOR FLUORIDE			
Test	Target/UOM	Range	Result
Actual pH of product	5.300 PH	5.000 - 5.500	5.251
Sample ID Tested 05/02/23 Tester RSHE Recorded 05/02/23			
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID Tested 05/02/23 Tester RSHE Recorded 05/02/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID Tested 05/02/23 Tester RSHE Recorded 05/02/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		T-23,GC,M25,M103,T-212,M90, M79
Sample ID Tested 05/02/23 Tester RSHE Recorded 05/02/23			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		05/30/24
Sample ID Tested 05/02/23 Tester RSHE Recorded 05/02/23			
SHELF LIFE OF PRODUCT	SHELF LIFE TIME		1 YEAR
Sample ID Tested 05/02/23 Tester RSHE Recorded 05/02/23			
FLUORIDE LEVEL	0.2 PPM	0.0 - 0.5	0.0
Sample ID Tested 05/02/23 Tester RSHE Recorded 05/02/23			
Lot #	3042875		
Made	05/02/23		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 05/02/23 at 12:55 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager