

Page: 1


Date: 05/03/23 at 2:17 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: ERL125F	7		
	ITRATE 0.1 NORMAL IN ISOPROPANOL		
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 05/03/23			
Tester DTRA			
Recorded 05/03/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 05/03/23			
Tester DTRA			
Recorded 05/03/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		05/03/23
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 05/03/23			
Tester DTRA			
Recorded 05/03/23			
MBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	20.0
SAMPLING DEGC	DEG C		
Sample ID			
Tested 05/03/23			
Tester DTRA			
Recorded 05/03/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		05/03/23
	MM/DD/YYYY		
Sample ID			
Tested 05/03/23			
Tester DTRA			
Recorded 05/03/23			
MBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	20.0
EG C	DEG C		
Sample ID			
Tested 05/03/23			
Tester DTRA			
Recorded 05/03/23			
Actual Normality	0.1000	0.0995 - 0.1005	0.1001
	EQ/L	2.5222 0000	0001
Sample ID			
Tested 05/03/23			
Tester DTRA			
Recorded 05/03/23			
CSP53: SILVER NITRATE STD ON	YES		YES
UTO-TITRA	METHOD		123
	IVIL IT IOD		
Sample ID			
Tested 05/03/23			
Tester DTRA			
Recorded 05/03/23			



Page: 2

Date: 05/03/23 at 2:17 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

		Cus	tomer PO # :			
N.I.S.T. Traceable to SF	RM 919 NaCl	YES YES		YES		
Sample ID						
Tested	05/03/23					
Tester	DTRA					
Recorded	05/03/23					
PREPARED TO ASTM D		YES		YES		
		YES				
Sample ID						
Tested	05/03/23					
Tester	DTRA					
Recorded	05/03/23					
Actual Density	00/00/20	0.82606	0.82377 - 0.82832	0.82674		
Actual Belisity		GM/ML	0.02077 0.02002	0.02074		
Sample ID		GIVI/IVIE				
Tested	05/03/23					
Tester	DTRA					
Recorded	05/03/23	CLEAR LIQUID		PASS		
APPEARANCE AND CO	LOR			PASS		
0 1 10		APPEARANCE				
Sample ID						
Tested	05/03/23					
Tester	DTRA					
Recorded	05/03/23					
INSTRUMENTS USED D	DURING	INSTRUMENT		M55,TID41094,BID52		
PREPARATION						
Sample ID						
Tested	05/03/23					
Tester	DTRA					
Recorded	05/03/23					
LOT NUMBER OF ORIG	INAL	LOT #		2111807		
MATERIAL						
Sample ID						
Tested	05/03/23					
Tester	DTRA					
Recorded	05/03/23					
EXPIRATION DATE		MM/DD/YY		05/16/23		
		MM/DD/YYYY				
Sample ID						
Tested	05/03/23					
Tester	DTRA					
Recorded	05/03/23					
SHELF LIFE OF PRODU		1 WEEK		1 WEEK		
		TIME				
Sample ID						
Tested	05/03/23					
Tester	DTRA					
Recorded	05/03/23					
Made 05/03	0/23					



Page: 3

Date: 05/03/23 at 2:17 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager