

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> TEN033F SULFIDE STANDARD 2000 PPM			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON</b>	YES		YES
<b>FILE</b>	YES		
Sample ID			
Tested 05/01/23			
Tester JENC			
Recorded 05/01/23			
<b>PPM SULFIDE</b>	2000 PPM	1980 - 2020	2006
Sample ID			
Tested 05/01/23			
Tester JENC			
Recorded 05/01/23			
<b>QCSP56:STANDARDIZATION</b>	YES		YES
<b>SODIUM SULFIDE SO</b>	QCSP56		
Sample ID			
Tested 05/01/23			
Tester JENC			
Recorded 05/01/23			
<b>Actual Density</b>	1.00893 GM/ML	1.00719 - 1.01069	1.00910
Sample ID			
Tested 05/01/23			
Tester JENC			
Recorded 05/01/23			
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 05/01/23			
Tester JENC			
Recorded 05/01/23			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M108,M100,VF1-Z33,M55,BID0 2253
Sample ID			
Tested 05/01/23			
Tester JENC			
Recorded 05/01/23			
<b>EXPIRATION DATE</b>	MM/DD/YY MM/DD/YYYY		06/06/23
Sample ID			
Tested 05/01/23			
Tester JENC			
Recorded 05/01/23			
<b>SHELF LIFE OF PRODUCT</b>	1 MONTH TIME		1 MONTH
Sample ID			
Tested 05/01/23			
Tester JENC			
Recorded 05/01/23			



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 2

Date: 05/01/23 at 9:09 AM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Lot #</b>	3042545
<b>Made</b>	05/01/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**