

Page: 1

Date: 05/04/23 at 1:27 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	BAS778F			
Hydrazine Buffer Solution pH 6.0 0.2 uM Filtered				
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	05/04/23			
Tester	BPHI			
Recorded	05/04/23			
PH OF PRODUCT/SOLUT	ION	6.000	5.980 - 6.020	5.985
		PH		
Sample ID				
-	05/04/23			
	BPHI			
	05/04/23			
CALCULATED MOLARITY		0.0050	0.0049 - 0.0051	0.0050
		MOLARITY (M/L)		
Sample ID		=		
	05/04/23			
	03/04/23 BPHI			
	05/04/23			
APPEARANCE AND COLO		CLEAR LIQUID		PASS
AFFEARANCE AND COL	On	APPEARANCE		1 700
0 1 15		AFFEARANCE		
Sample ID	0=104100			
	05/04/23			
	BPHI			
	05/04/23	OLEAR LIQUID		MO4 T 40 M400
INSTRUMENTS USED DU	IRING	CLEAR LIQUID		M91,T-16,M106
PREPARATION		INSTRUMENT		
Sample ID				
	05/04/23			
	BPHI			
	05/04/23			
EXPIRATION DATE		MM/DD/YYYY		02/28/24
		MM/DD/YYYY		
Sample ID				
Tested	05/04/23			
Tester	BPHI			
Recorded	05/04/23			
Lot # 305031	.0			
Made 05/04/				
wade 03/04/	۷ ک			



Page: 2

Date: 05/04/23 at 1:27 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager