

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: ERL213F			
POTASSIUM HYDROXIDE 0.005 NORMAL IN METHANOL (1LG)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 05/03/23			
Tester CCOA			
Recorded 05/03/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 05/03/23			
Tester CCOA			
Recorded 05/03/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		05/03/23
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 05/03/23			
Tester CCOA			
Recorded 05/03/23			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	20.9
SAMPLING DEGC	DEG C		
Sample ID			
Tested 05/03/23			
Tester CCOA			
Recorded 05/03/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		05/03/23
	MM/DD/YYYY		
Sample ID			
Tested 05/03/23			
Tester CCOA			
Recorded 05/03/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	20.9
DEG C	DEG C		
Sample ID			
Tested 05/03/23			
Tester CCOA			
Recorded 05/03/23			
NORMALITY TO 4 DEC PLACES	0.0050	0.0043 - 0.0055	0.0050
	EQUIV/LITER		
Sample ID			
Tested 05/03/23			
Tester CCOA			
Recorded 05/03/23			
Actual Density	0.79257	0.79116 - 0.79399	0.79224
	GM/ML		
Sample ID			
Tested 05/03/23			
Tester CCOA			
Recorded 05/03/23			

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 84 (KHP)	YES YES	YES
Sample ID		
Tested	05/03/23	
Tester	CCOA	
Recorded	05/03/23	
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested	05/03/23	
Tester	CCOA	
Recorded	05/03/23	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M81,VF4-N21,M55,BID03828,P IDP20-73
Sample ID		
Tested	05/03/23	
Tester	CCOA	
Recorded	05/03/23	
SHELF LIFE OF PRODUCT	1 MONTH TIME	1 MONTH
Sample ID		
Tested	05/03/23	
Tester	CCOA	
Recorded	05/03/23	
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY	06/15/23
Sample ID		
Tested	05/03/23	
Tester	CCOA	
Recorded	05/03/23	
Lot #	3041022	
Made	05/03/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.


Quality Manager