

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product	FIS101F			
5		Standard 90 ppm Certified		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	05/02/23			
Tester	MHUS			
Recorded	05/02/23			
ALKALINITY AS CACO	3	90.0	85.0 - 95.0	89.8
		PPM		
Sample ID				
Tested	05/02/23			
Tester	MHUS			
Recorded	05/02/23			
APPEARANCE AND CO	LOR	CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	05/02/23			
Tester	MHUS			
Recorded	05/02/23			
INSTRUMENTS USED	DURING	INSTRUMENT		M91,VF2-A7,M55,BURET:3902
PREPARATION				
Sample ID				
Tested	05/02/23			
Tester	MHUS			
Recorded	05/02/23			
EXPIRATION DATE		MM/DD/YYYY		09/30/23
Sample ID				
Tested	05/02/23			
Tester	MHUS			
Recorded	05/02/23			
Lot # 30426	533			
Made 05/02				

ichard St

Quality Manager