

Page: 1

Date: 05/02/23 at 8:36 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

	045			
Packaged Product				
	IODINE	SOLUTION 0.006 NORMAL		
Test		Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	05/02/23			
Tester	JENC			
Recorded	05/02/23			
NORMALITY TO 4 DEC PLACES		0.0060	0.0057 - 0.0063	0.0057
		EQUIV/LITER		
Sample ID				
Tested	05/02/23			
Tester	JENC			
Recorded	05/02/23			
TRACEABLE TO NIST	SRM 136	YES		YES
		YES		
Sample ID				
Tested	05/02/23			
Tester	JENC			
Recorded	05/02/23			
APPEARANCE AND CO		CLEAR BROWN LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	05/02/23			
Tester	JENC			
Recorded	05/02/23			
INSTRUMENTS USED	DURING	INSTRUMENT		P.113,P51,P389,VF3-J11
PREPARATION				
Sample ID				
Tested	05/02/23			
Tester	JENC			
Recorded	05/02/23			
SHELF LIFE OF PROD	UCT	1 YEAR		1 YEAR
		TIME		
Sample ID				
Tested	05/02/23			
Tester	JENC			
Recorded	05/02/23			
EXPIRATION DATE		MM/DD/YYYY		05/30/24
		MM/DD/YYYY		
Sample ID				
Tested	05/02/23			
Tester	JENC			
Recorded	05/02/23			
Lot # 3042	745			
Made 05/0				
Iviaue 00/0	2,29			



Page: 2

Date: 05/02/23 at 8:36 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager