

Made 05/09/23

Page: 1

Date: 05/09/23 at 11:23 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packag	ed Product:	ERL254F			
. uonug	ca i ioaaot		NITROGEN 10 PPM AS "N"	SMART. NH4 TEST	
Test			Target/UOM	Range	Result
	ED TO FORMU	LATION ON	YES	90	YES
FILE			YES		
	Sample ID				
	Tested	05/09/23			
	Tester	MMAT			
	Recorded	05/09/23			
PPM NITROGEN 2 DECIMAL		IMAL	10.00	9.50 - 10.50	10.21
PLACES			PPM		
	Sample ID				
	Tested	05/09/23			
	Tester	MMAT			
	Recorded	05/09/23			
APPEARANCE AND COLOR		LOR	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	05/09/23			
	Tester	MMAT			
	Recorded	05/09/23			
INSTRUM	MENTS USED D		INSTRUMENT		P90,VF1-Z32,M107
PREPAR	ATION				, ,
	Sample ID				
	Tested	05/09/23			
	Tester	MMAT			
	Recorded	05/09/23			
SHELF LIFE OF PRODUCT			2 MONTHS		2 MONTHS
			TIME		
	Sample ID				
	Tested	05/09/23			
	Tester	MMAT			
	Recorded	05/09/23			
EXPIRATION DATE		00/00/20	MM/DD/YYYY		07/14/23
			MM/DD/YYYY		2
	Sample ID		= =		
	Tested	05/09/23			
	Tester	MMAT			
	Recorded	05/09/23			
	ot # 30509				
L	OL# 30303				



Page: 2

Date: 05/09/23 at 11:23 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager