

Page: 1

Date: 05/10/23 at 6:23 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	: TEN003F			
	BUFFER 1	REFERENCE pH 2.0		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	05/04/23			
Tester	JRIO			
Recorded	05/04/23			
PH OF PRODUCT/SOLUTION		2.000	1.980 - 2.020	2.007
		PH		
Sample ID				
Tested	05/04/23			
Tester	JRIO			
Recorded	05/04/23			
QCSP 20 PH CALIBRA		YES		YES
METHODS		METHOD		
Sample ID				
Tested	05/04/23			
Tester	JRIO			
Recorded	05/04/23			
APPEARANCE AND CO		ORANGE LIQUID		PASS
		APPEARANCE		
Sample ID		711 2711 01102		
Tested	05/04/23			
Tester	JRIO			
Recorded	05/04/23			
INSTRUMENTS USED I		INSTRUMENT		M115,GC,VF12-95
PREPARATION		INOTTONENT		18/11/0,00, \$1 12 33
Sample ID Tested	05/04/00			
Tested	05/04/23 JRIO			
Recorded	05/04/23			
EXPIRATION DATE	03/04/23	MM/DD/YYYY		05/30/24
EAFINATION DATE		MM/DD/YYYY		03/30/24
Sample ID		, 22,		
Tested	05/10/23			
Tester	JRIO			
Recorded	05/10/23			
Lot # 30428				
Made 05/04/23				
wade 00/0	1,40			



Page: 2

Date: 05/10/23 at 6:23 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager