

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	9141		
	TAN STANDARD 0.1 MG/G KOH	IOTAL ACID NUMBER STANDARD	
Test	Target/UOM	Range	Result
PREPARED TO FORMU		~	YES
FILE	YES		
Sample ID			
Tested	05/10/23		
Tester	AROS		
Recorded	05/10/23		
TOTAL ACID NUMBER (TAN) VALUE 0.10	0.08 - 0.12	0.10
MG KOH/G	MGM KOH/GM		
Sample ID			
Tested	05/10/23		
Tester	AROS		
Recorded	05/10/23		
APPEARANCE AND COL	LOR CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested	05/10/23		
Tester	AROS		
Recorded	05/10/23		
INSTRUMENTS USED D	URING INSTRUMENT		M100
PREPARATION			
Sample ID			
Tested	05/10/23		
Tester	AROS		
Recorded	05/10/23		
Lot # 3050945			
Made 05/10	/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

hichard Stot

Quality Manager