

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> OXY097F DNTB REAGENT FOR SULFITE ANALYZER			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES YES		YES
Sample ID			
Tested 05/10/23			
Tester BPHI			
Recorded 05/10/23			
<b>PH OF PRODUCT/SOLUTION</b>	6.900 PH	6.880 - 6.920	6.900
Sample ID			
Tested 05/10/23			
Tester BPHI			
Recorded 05/10/23			
<b>CALCULATED GPL DNTB</b>	4.0000 GPL	3.9995 - 4.0005	4.0000
Sample ID			
Tested 05/10/23			
Tester BPHI			
Recorded 05/10/23			
<b>APPEARANCE AND COLOR</b>	CLEAR PALE YELLOW TO LIME GREEN APPEARANCE		PASS
Sample ID			
Tested 05/10/23			
Tester BPHI			
Recorded 05/10/23			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M103,VF12-93,M106,PID41,PI D22
Sample ID			
Tested 05/10/23			
Tester BPHI			
Recorded 05/10/23			
<b>Actual Density</b>	1.00421 GM/ML	1.00198 - 1.00645	1.00527
Sample ID			
Tested 05/10/23			
Tester BPHI			
Recorded 05/10/23			
<b>EXPIRATION DATE</b>	MM/DD/YYYY MM/DD/YYYY		05/30/24
Sample ID			
Tested 05/10/23			
Tester BPHI			
Recorded 05/10/23			
<b>Lot #</b>	3050207		
<b>Made</b>	05/10/23		



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 2

Date: 05/10/23 at 9:53 AM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**