



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 8430			
SILVER NITRATE 0.1 NORMAL NIST TRACEABLE SOLUTION			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 05/11/23			
Tester DTRA			
Recorded 05/11/23			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 05/11/23			
Tester DTRA			
Recorded 05/11/23			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		05/11/23
Sample ID			
Tested 05/11/23			
Tester DTRA			
Recorded 05/11/23			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0	10.0 - 35.0	20.0
Sample ID			
Tested 05/11/23			
Tester DTRA			
Recorded 05/11/23			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		05/11/23
Sample ID			
Tested 05/11/23			
Tester DTRA			
Recorded 05/11/23			
<b>AMBIENT TEST TEMPERATURE DEGC</b>	20.0	15.0 - 30.0	20.0
Sample ID			
Tested 05/11/23			
Tester DTRA			
Recorded 05/11/23			
<b>Actual Normality</b>	0.1000	0.0995 - 0.1005	0.1002
	EQ/L		
Sample ID			
Tested 05/11/23			
Tester DTRA			
Recorded 05/11/23			
<b>QCSP53: SILVER NITRATE STD ON AUTO-TITRA</b>	YES		YES
	METHOD		
Sample ID			
Tested 05/11/23			
Tester DTRA			
Recorded 05/11/23			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>Actual Density</b>	1.01229 GM/ML	1.00868 - 1.01590	1.01239
Sample ID			
Tested	05/11/23		
Tester	DTRA		
Recorded	05/11/23		
<b>VERIFICATION BY QCSP78</b>	PASS PASS		PASS
Sample ID			
Tested	05/11/23		
Tester	DTRA		
Recorded	05/11/23		
<b>N.I.S.T. Traceable to SRM 919 NaCl</b>	YES YES		YES
Sample ID			
Tested	05/11/23		
Tester	DTRA		
Recorded	05/11/23		
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	05/11/23		
Tester	DTRA		
Recorded	05/11/23		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M113,T261,M55
Sample ID			
Tested	05/11/23		
Tester	DTRA		
Recorded	05/11/23		
<b>EXPIRATION DATE</b>	MM/DD/YY MM/DD/YYYY		05/30/25
Sample ID			
Tested	05/11/23		
Tester	DTRA		
Recorded	05/11/23		
<b>NORMALITY USING AUTO-TITRATOR</b>	0.1000 EQUIV/LITER	0.0985 - 0.1015	0.1002
Sample ID			
Tested	05/11/23		
Tester	DTRA		
Recorded	05/11/23		
<b>SHELF LIFE OF PRODUCT</b>	SHELF LIFE TIME		2 YEARS
Sample ID			
Tested	05/11/23		
Tester	DTRA		
Recorded	05/11/23		

**Lot #** 3050835

**Made** 05/11/23



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 3

Date: 05/11/23 at 1:39 PM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**