



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8580			
SODIUM CARBONATE 0.05 NORMAL			
Test	Target/UOM	Range	Result
STANDARDIZED USING QCSP36	YES		YES
	PASS		
Sample ID			
Tested 05/11/23			
Tester MHUS			
Recorded 05/11/23			
N.I.S.T. Traceable to SRM 723	Yes		YES
	YES		
Sample ID			
Tested 05/11/23			
Tester MHUS			
Recorded 05/11/23			
Actual Density	1.00097	1.00007 - 1.00187	1.00102
	GM/ML		
Sample ID			
Tested 05/11/23			
Tester MHUS			
Recorded 05/11/23			
PREPARED TO FORMULATION ON FILE	YES		YES
	YES		
Sample ID			
Tested 05/11/23			
Tester MHUS			
Recorded 05/11/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 05/11/23			
Tester MHUS			
Recorded 05/11/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M112,VF12-82,BID-04099,PID-P120
Sample ID			
Tested 05/11/23			
Tester MHUS			
Recorded 05/11/23			
EXPIRATION DATE	MM/DD/YY		05/30/24
	MM/DD/YYYY		
Sample ID			
Tested 05/11/23			
Tester MHUS			
Recorded 05/11/23			
SHELF LIFE OF PRODUCT	1 YEAR		1 YEAR
	TIME		
Sample ID			
Tested 05/11/23			
Tester MHUS			
Recorded 05/11/23			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results

Customer PO # :

Actual Normality	0.0500	0.0495 - 0.0505	0.0500
	EQ/L		
Sample ID			
Tested	05/15/23		
Tester	MHUS		
Recorded	05/15/23		
Lot #	3050959		
Made	05/11/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager