

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	6945			
3	CONDUCI	IVITY STANDARD 50,000 u	MHOS	
Test		Target/UOM	Range	Result
PREPARED TO FORMULA	ATION ON	YES		YES
FILE		YES		
Sample ID				
	05/16/23			
	RSHE			
)5/16/23			
CONDUCTIVITY @25 C.		50000 US/CM	49999 - 50001	50000
Sample ID				
)5/16/23			
	RSHE			
Recorded 0)5/16/23			
QCSP30:CONDUCTIVITY		YES		YES
STD&MEASUREMENT		QCSP30		
Sample ID				
)5/16/23			
Tester F	RSHE			
Recorded 0)5/16/23			
TRACEABLE TO N.I.S.T. S	SRM 3191	YES		YES
		YES		
Sample ID				
Tested 0	05/16/23			
Tester F	RSHE			
Recorded 0)5/16/23			
APPEARANCE AND COLO	DR	CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested 0)5/16/23			
Tester F	RSHE			
	05/16/23			
INSTRUMENTS USED DUI	RING	INSTRUMENT		T-229,M92
PREPARATION				
Sample ID				
	05/16/23			
	RSHE			
	05/16/23			
EXPIRATION DATE		MM/DD/YY		05/30/24
		MM/DD/YYYY		
Sample ID				
	05/16/23			
	RSHE			
)5/16/23 -			
SHELF LIFE OF PRODUC	I	1 YEAR		1 YEAR
		TIME		
Sample ID				
)5/16/23			
	RSHE			
Recorded 0)5/16/23			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	3051578
Made	05/16/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

hard

Quality Manager