



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 9106			
SULFURIC ACID 10.0 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 05/19/23			
Tester CCOA			
Recorded 05/19/23			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 05/19/23			
Tester CCOA			
Recorded 05/19/23			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		05/19/23
Sample ID			
Tested 05/19/23			
Tester CCOA			
Recorded 05/19/23			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	24.3
Sample ID			
Tested 05/19/23			
Tester CCOA			
Recorded 05/19/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		05/19/23
Sample ID			
Tested 05/19/23			
Tester CCOA			
Recorded 05/19/23			
AMBIENT TEST TEMPERATURE DEGC	20.0 DEG C	15.0 - 30.0	20.9
Sample ID			
Tested 05/19/23			
Tester CCOA			
Recorded 05/19/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested 05/19/23			
Tester CCOA			
Recorded 05/19/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		T49,T199,M55,P111,BID00488, TID4108
Sample ID			
Tested 05/19/23			
Tester CCOA			
Recorded 05/19/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

NORMALITY	10.000	9.950 - 10.050	10.004
	EQUIV/LITER		
Sample ID			
Tested	05/19/23		
Tester	CCOA		
Recorded	05/19/23		
N.I.S.T. Traceable to SRM 84	YES		YES
	YES		
Sample ID			
Tested	05/19/23		
Tester	CCOA		
Recorded	05/19/23		
EXPIRATION DATE	MM/DD/YY		05/30/25
	MM/DD/YYYY		
Sample ID			
Tested	05/19/23		
Tester	CCOA		
Recorded	05/19/23		
Actual Density	1.28616	1.28298 - 1.28934	1.28593
	GM/ML		
Sample ID			
Tested	05/19/23		
Tester	CCOA		
Recorded	05/19/23		
Lot #	3051548		
Made	05/19/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager