



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8820			
SODIUM THIOSULFATE 1.0 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	1.0000 EQ/L	0.9990 - 1.0010	1.0001
Sample ID			
Tested	05/22/23		
Tester	DTRA		
Recorded	05/22/23		
QCSP10:STANDARDIZATION OF SODIUM THIOSUL	YES QCSP10		YES
Sample ID			
Tested	05/22/23		
Tester	DTRA		
Recorded	05/22/23		
TITRATION BLANK	0.01 MLS	0.00 - 0.02	0.02
Sample ID			
Tested	05/22/23		
Tester	DTRA		
Recorded	05/22/23		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	05/22/23		
Tester	DTRA		
Recorded	05/22/23		
Actual Density	1.11980 GM/ML	1.11845 - 1.12115	1.12032
Sample ID			
Tested	05/22/23		
Tester	DTRA		
Recorded	05/22/23		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	05/22/23		
Tester	DTRA		
Recorded	05/22/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M112,T-25BID-00787,PID-M81
Sample ID			
Tested	05/22/23		
Tester	DTRA		
Recorded	05/22/23		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		09/30/23
Sample ID			
Tested	05/22/23		
Tester	DTRA		
Recorded	05/22/23		



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT

4 MONTHS
 TIME

4 MONTHS

Sample ID

Tested 05/22/23

Tester DTRA

Recorded 05/22/23

Lot # 3051706

Made 05/22/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager