

Page: 1

Date: 06/01/23 at 1:07 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	A1126F			
_		ACID 0.1 NORMAL		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES	<u> </u>	YES
FILE		YES		
Sample ID				
Tested	06/01/23			
Tester	CCOA			
Recorded	06/01/23			
Actual Normality		0.1000	0.0990 - 0.1010	0.1002
•		EQ/L		
Sample ID				
Tested	06/01/23			
Tester	CCOA			
Recorded	06/01/23			
QCSP37:STANDARDIZA		YES		YES
ACETIC ACID SOLN		QCSP37		120
Sample ID		Q001 07		
Tested	00/04/00			
	06/01/23			
Tester	CCOA			
Recorded	06/01/23	CLEAR LIQUID		DACC
APPEARANCE AND COL	LOR	CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	06/01/23			
Tester	CCOA			
Recorded	06/01/23			
INSTRUMENTS USED DURING		INSTRUMENT		M103,VF2-B3
PREPARATION				
Sample ID				
Tested	06/01/23			
Tester	CCOA			
Recorded	06/01/23			
Actual Density		0.99910	0.99902 - 0.99918	0.99915
-		GM/ML		
Sample ID				
Tested	06/01/23			
Tester	CCOA			
Recorded	06/01/23			
EXPIRATION DATE	-	MM/DD/YY		06/30/24
		MM/DD/YYYY		
Sample ID				
Tested	06/01/23			
Tester	CCOA			
Recorded	06/01/23			
Lot # 3053160				
Made 06/01	/ 23			



Page: 2

Date: 06/01/23 at 1:07 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager