



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8800			
SODIUM THIOSULFATE 0.025 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.0250	0.0245 - 0.0255	0.0249
	EQ/L		
Sample ID			
Tested	06/02/23		
Tester	MHUS		
Recorded	06/02/23		
QCSP10:STANDARDIZATION OF SODIUM THIOSUL	YES		YES
	QCSP10		
Sample ID			
Tested	06/02/23		
Tester	MHUS		
Recorded	06/02/23		
PREPARED TO FORMULATION ON FILE	YES		YES
	YES		
Sample ID			
Tested	06/02/23		
Tester	MHUS		
Recorded	06/02/23		
Actual Density	1.00161	1.00135 - 1.00187	1.00167
	GM/ML		
Sample ID			
Tested	06/02/23		
Tester	MHUS		
Recorded	06/02/23		
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested	06/02/23		
Tester	MHUS		
Recorded	06/02/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M112,T4
Sample ID			
Tested	06/02/23		
Tester	MHUS		
Recorded	06/02/23		
TRACEABLE TO NIST SRM 136	YES		YES
	YES		
Sample ID			
Tested	06/02/23		
Tester	MHUS		
Recorded	06/02/23		
EXPIRATION DATE	MM/DD/YY		12/30/23
	MM/DD/YYYY		
Sample ID			
Tested	06/02/23		
Tester	MHUS		
Recorded	06/02/23		



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT

6 MONTHS
 TIME

6 MONTHS

Sample ID

Tested 06/02/23

Tester MHUS

Recorded 06/02/23

Lot # 3060102

Made 06/02/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager