

Page: 1

Date: 06/02/23 at 3:37 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results
Customer PO #:

			Ouston	ilei FO#.	
Packag	ed Product:	1670			
·			M CHLORIDE Solution 2.75	% w/v APHA for BOD	
Test			Target/UOM	Range	Result
	ED TO FORMU	LATION ON	YES	<u> </u>	YES
FILE			YES		
	Sample ID				
	Tested	06/02/23			
	Tester	JENC			
	Recorded	06/02/23			
APPEARANCE AND COLOR		LOR	CLEAR WATER WHITE		PASS
			LIQUID		
			APPEARANCE		
	Sample ID				
	Tested	06/02/23			
	Tester	JENC			
	Recorded	06/02/23			
Actual D	ensity		1.02113	1.02024 - 1.02202	1.02093
	-		GM/ML		
	Sample ID				
	Tested	06/02/23			
	Tester	JENC			
	Recorded	06/02/23			
INSTRUMENTS USED DURING		URING	LAB EQUIPMENT		M112,T165
PREPARATION			INSTRUMENT		
	Sample ID				
	Tested	06/02/23			
	Tester	JENC			
	Recorded	06/02/23			
W/V% CACL2			2.75	2.73 - 2.77	2.74
			% CACL2		
	Sample ID				
	Tested	06/02/23			
	Tester	JENC			
	Recorded	06/02/23			
EXPIRAT	ION DATE		MM/DD/YY		06/30/24
			MM/DD/YYYY		
	Sample ID				
	Tested	06/02/23			
	Tester	JENC			
	Recorded	06/02/23			
1	ot # 30602				
	ade 06/02				
IVI	uuc °°, °2	, = =			



Page: 2

Date: 06/02/23 at 3:37 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager