

Page: 1

Date: 06/02/23 at 4:13 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: 6	924		
CONDUCTIVITY STANDARD 2,000 uMHOS			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATIO			YES
FILE	YES		
Sample ID			
Tested 06/02			
Tester BPH			
Recorded 06/02			
CONDUCTIVITY @25 C.	2000	1999 - 2001	2000
	US/CM		
Sample ID			
Tested 06/02	2/23		
Tester BPH			
Recorded 06/02	2/23		
CSP30:CONDUCTIVITY	YES		YES
STD&MEASUREMENT	QCSP30		
Sample ID			
Tested 06/02	2/23		
Tester BPH			
Recorded 06/02	2/23		
RACEABLE TO N.I.S.T. SRM	3191 YES		YES
	YES		
Sample ID			
Tested 06/02	2/23		
Tester BPH			
Recorded 06/02	2/23		
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 06/02	2/23		
Tester BPHI			
Recorded 06/02			
EXPIRATION DATE	MM/DD/YYYY		06/30/24
	MM/DD/YYYY		
Sample ID			
Tested 06/02	2/23		
Tester BPH			
Recorded 06/02			
NSTRUMENTS USED DURING	G INSTRUMENT		M92,T-229
REPARATION			
Sample ID			
Tested 06/02	2/23		
Tester BPHI			
Recorded 06/02			
Lot # 3060230			
Made 06/02/23			



Page: 2

Date: 06/02/23 at 4:13 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager