



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 9098			
SULFURIC ACID 5.25 NORMAL NIST TRACEABLE SOLUTION			
Test	Target/UOM	Range	Result
N.I.S.T. Traceable to SRM 723			
	Yes		YES
	YES		
Sample ID			
Tested	06/07/23		
Tester	JENC		
Recorded	06/07/23		
SAMPLE COLLECTION & HANDLING			
	YES		YES
PER LAB040			
	YES		
Sample ID			
Tested	06/07/23		
Tester	JENC		
Recorded	06/07/23		
DATE OF SAMPLING FOR TEST			
	MM/DD/YYYY		06/07/23
PURPOSES			
	MM/DD/YYYY		
Sample ID			
Tested	06/07/23		
Tester	JENC		
Recorded	06/07/23		
AMBIENT TEMPERATURE DURING			
	25.0	10.0 - 35.0	21.6
SAMPLING DEGC			
	DEG C		
Sample ID			
Tested	06/07/23		
Tester	JENC		
Recorded	06/07/23		
DATE OF TESTING OF SAMPLE			
	MM/DD/YYYY		06/07/23
	MM/DD/YYYY		
Sample ID			
Tested	06/07/23		
Tester	JENC		
Recorded	06/07/23		
AMBIENT TEST TEMPERATURE			
	20.0	15.0 - 30.0	21.6
DEG C			
	DEG C		
Sample ID			
Tested	06/07/23		
Tester	JENC		
Recorded	06/07/23		
QCSP5: STANDARDIZATION OF			
	YES		YES
AQUEOUS ACIDS			
	METHOD		
Sample ID			
Tested	06/07/23		
Tester	JENC		
Recorded	06/07/23		
Actual Normality			
	5.2500	5.2400 - 5.2600	5.2527
	EQ/L		
Sample ID			
Tested	06/07/23		
Tester	JENC		
Recorded	06/07/23		



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	06/07/23		
Tester	JENC		
Recorded	06/07/23		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	06/07/23		
Tester	JENC		
Recorded	06/07/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		GC,VF2-B19,M55,BURET:0225 3
Sample ID			
Tested	06/07/23		
Tester	JENC		
Recorded	06/07/23		
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		06/30/25
Sample ID			
Tested	06/07/23		
Tester	JENC		
Recorded	06/07/23		
Actual Density	1.15678 GM/ML	1.15583 - 1.15773	1.15712
Sample ID			
Tested	06/07/23		
Tester	JENC		
Recorded	06/07/23		
Lot #	3060738		
Made	06/07/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager