



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

Test	Target/UOM	Range	Result
<b>Packaged Product:</b> ISP125F			
	HYDROCHLORIC ACID 0.01 NORMAL IN IPA		
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 06/08/23			
Tester JENC			
Recorded 06/08/23			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 06/08/23			
Tester JENC			
Recorded 06/08/23			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		06/08/23
Sample ID			
Tested 06/08/23			
Tester JENC			
Recorded 06/08/23			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0 DEG C	10.0 - 35.0	22.1
Sample ID			
Tested 06/08/23			
Tester JENC			
Recorded 06/08/23			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY MM/DD/YYYY		06/08/23
Sample ID			
Tested 06/08/23			
Tester JENC			
Recorded 06/08/23			
<b>AMBIENT TEST TEMPERATURE DEGC</b>	20.0 DEG C	15.0 - 30.0	22.1
Sample ID			
Tested 06/08/23			
Tester JENC			
Recorded 06/08/23			
<b>NORMALITY TO 4 DEC PLACES</b>	0.0100 EQUIV/LITER	0.0095 - 0.0105	0.0097
Sample ID			
Tested 06/08/23			
Tester JENC			
Recorded 06/08/23			
<b>TRACEABLE TO N.I.S.T. SRM 84</b>	YES YES		YES
Sample ID			
Tested 06/08/23			
Tester JENC			
Recorded 06/08/23			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	06/08/23		
Tester	JENC		
Recorded	06/08/23		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		GC,VF4-N6
Sample ID			
Tested	06/08/23		
Tester	JENC		
Recorded	06/08/23		
<b>SHELF LIFE OF PRODUCT</b>	6 MONTHS TIME		6 MONTHS
Sample ID			
Tested	06/08/23		
Tester	JENC		
Recorded	06/08/23		
<b>EXPIRATION DATE</b>	MM/DD/YYYY MM/DD/YYYY		12/30/23
Sample ID			
Tested	06/08/23		
Tester	JENC		
Recorded	06/08/23		
<b>Actual Density</b>	0.78587 GM/ML	0.78546 - 0.78620	0.78610
Sample ID			
Tested	06/08/23		
Tester	JENC		
Recorded	06/08/23		
<b>Lot #</b>	3060840		
<b>Made</b>	06/08/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**