

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: UC115F SODIUM HYDROXIDE 0.5 NORMAL STD ALCOHOLIC-8647			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 06/07/23			
Tester MHUS			
Recorded 06/07/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 06/07/23			
Tester MHUS			
Recorded 06/07/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		06/07/23
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 06/07/23			
Tester MHUS			
Recorded 06/07/23			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	22.1
SAMPLING DEGC	DEG C		
Sample ID			
Tested 06/07/23			
Tester MHUS			
Recorded 06/07/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		06/07/23
	MM/DD/YYYY		
Sample ID			
Tested 06/07/23			
Tester MHUS			
Recorded 06/07/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	22.1
DEG C	DEG C		
Sample ID			
Tested 06/07/23			
Tester MHUS			
Recorded 06/07/23			
Actual Normality	0.5000	0.4995 - 0.5005	0.4997
	EQ/L		
Sample ID			
Tested 06/07/23			
Tester MHUS			
Recorded 06/07/23			
QCSP7: STANDARDIZATION OF	YES		YES
AQUEOUS BASES	METHOD		
Sample ID			
Tested 06/07/23			
Tester MHUS			
Recorded 06/07/23			

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 84 (KHP)		YES	YES
		YES	
Sample ID			
Tested 06/07/23			
Tester MHUS			
Recorded 06/07/23			
Actual Density	0.89612	0.89396 - 0.89829	0.89429
	GM/ML		
Sample ID			
Tested 06/07/23			
Tester MHUS			
Recorded 06/07/23			
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested 06/07/23			
Tester MHUS			
Recorded 06/07/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M55,M112,GC,EF2-JC9,VF5-E 3	
Sample ID			
Tested 06/07/23			
Tester MHUS			
Recorded 06/07/23			
SHELF LIFE OF PRODUCT	6 MONTHS		6 MONTHS
	TIME		
Sample ID			
Tested 06/07/23			
Tester MHUS			
Recorded 06/07/23			
EXPIRATION DATE	MM/DD/YYYY		12/30/23
	MM/DD/YYYY		
Sample ID			
Tested 06/07/23			
Tester MHUS			
Recorded 06/07/23			
Lot #	3060716		
Made	06/07/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 06/07/23 at 5:02 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager