



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: ERL114AF			
	POTASSIUM HYDROXIDE 0.1 NORMAL IN IPA		
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 06/06/23			
Tester JENC			
Recorded 06/06/23			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 06/06/23			
Tester JENC			
Recorded 06/06/23			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		06/06/23
Sample ID			
Tested 06/06/23			
Tester JENC			
Recorded 06/06/23			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	21.0
Sample ID			
Tested 06/06/23			
Tester JENC			
Recorded 06/06/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		06/06/23
Sample ID			
Tested 06/06/23			
Tester JENC			
Recorded 06/06/23			
AMBIENT TEST TEMPERATURE DEG C	20.0 DEG C	15.0 - 30.0	21.0
Sample ID			
Tested 06/06/23			
Tester JENC			
Recorded 06/06/23			
LOT NUMBER OF ORIGINAL MATERIAL	BULK LOT NUMBER LOT #		3042419
Sample ID			
Tested 06/06/23			
Tester JENC			
Recorded 06/06/23			
Actual Normality	0.1000 EQ/L	0.0995 - 0.1005	0.1004
Sample ID			
Tested 06/06/23			
Tester JENC			
Recorded 06/06/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

Actual Density	0.79116 GM/ML	0.78994 - 0.79237	0.79133
Sample ID			
Tested	06/06/23		
Tester	JENC		
Recorded	06/06/23		
N.I.S.T. Traceable to SRM 84 (KHP)	YES YES		YES
Sample ID			
Tested	06/06/23		
Tester	JENC		
Recorded	06/06/23		
APPEARANCE AND COLOR	CLEAR-SLIGHTLY HAZY LIQUID APPEARANCE		PASS
Sample ID			
Tested	06/06/23		
Tester	JENC		
Recorded	06/06/23		
INSTRUMENTS USED DURING PREPARATION	LAB EQUIPMENT INSTRUMENT		M55,M91,BID02253,TID4100
Sample ID			
Tested	06/06/23		
Tester	JENC		
Recorded	06/06/23		
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		06/20/23
Sample ID			
Tested	06/06/23		
Tester	JENC		
Recorded	06/06/23		
SHELF LIFE OF PRODUCT	1 WEEK TIME		1 WEEK
Sample ID			
Tested	06/06/23		
Tester	JENC		
Recorded	06/06/23		
Lot #	3060511		
Made	06/06/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 06/06/23 at 9:54 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager