

Page: 1

Date: 06/06/23 at 3:53 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	EQS148F			
	SULFURIO	C ACID 0.01 NORMAL		
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES	1131130	YES
FILE		YES		
Sample ID				
•	06/06/23			
	AFRE			
Recorded	06/06/23			
Actual Normality		0.0100	0.0095 - 0.0105	0.0100
		EQ/L		
Sample ID				
	06/06/23			
	AFRE			
	06/06/23			
APPEARANCE AND COL		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	06/06/23			
Tester	AFRE			
Recorded	06/06/23			
INSTRUMENTS USED DU	JRING	INSTRUMENT		P58,M100
PREPARATION				
Sample ID				
Tested	06/06/23			
Tester	AFRE			
Recorded	06/06/23			
EXPIRATION DATE		MM/DD/YY		06/30/24
		MM/DD/YYYY		
Sample ID				
Tested	06/06/23			
Tester	AFRE			
Recorded	06/06/23			
SHELF LIFE OF PRODUC	СТ	SHELF LIFE		1 YEAR
		TIME		
Sample ID				
Tested	06/06/23			
Tester	AFRE			
	06/06/23			
Actual Density		0.99863	0.99843 - 0.99883	0.99850
		GM/ML		
Sample ID				
	06/06/23			
	AFRE			
Recorded	06/06/23			
Lot # 3060643				
Made 06/06	/23			
WIGGE 55, 55,	-			



Page: 2

Date: 06/06/23 at 3:53 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager