



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: 1501			
	BUFFER REFERENCE pH 7.0/CLEAR		
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 06/06/23			
Tester BPHI			
Recorded 06/06/23			
PH OF PRODUCT/SOLUTION	7.000 PH	6.990 - 7.010	6.998
Sample ID			
Tested 06/06/23			
Tester BPHI			
Recorded 06/06/23			
QCSP 20 PH CALIBRATION METHODS	YES METHOD		YES
Sample ID			
Tested 06/06/23			
Tester BPHI			
Recorded 06/06/23			
Actual Temp	25.0 DEG C	24.0 - 26.0	25.0
Sample ID			
Tested 06/06/23			
Tester BPHI			
Recorded 06/06/23			
pH Traceable to N.I.S.T. SRM'S	YES YES		YES
Sample ID			
Tested 06/06/23			
Tester BPHI			
Recorded 06/06/23			
PH NIST TRACEABLE TO SRM 185	YES YES		YES
Sample ID			
Tested 06/06/23			
Tester BPHI			
Recorded 06/06/23			
NIST TRACEABLE TO SRM186I AND 186II	YES YES		YES
Sample ID			
Tested 06/06/23			
Tester BPHI			
Recorded 06/06/23			
PH NIST TRACEABLE TO SRM 187	YES YES		YES
Sample ID			
Tested 06/06/23			
Tester BPHI			
Recorded 06/06/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE	PASS
Sample ID		
Tested	06/06/23	
Tester	BPHI	
Recorded	06/06/23	
INSTRUMENTS USED DURING PREPARATION	LAB EQUIPMENT INSTRUMENT	M106,M87,T-IX
Sample ID		
Tested	06/06/23	
Tester	BPHI	
Recorded	06/06/23	
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	06/30/25
Sample ID		
Tested	06/06/23	
Tester	BPHI	
Recorded	06/06/23	
SHELF LIFE OF PRODUCT	SHELF LIFE TIME	2 YEARS
Sample ID		
Tested	06/06/23	
Tester	BPHI	
Recorded	06/06/23	
Lot #	3052502	
Made	06/06/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager