



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 4180			
HYDROCHLORIC ACID 0.4 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.4000 EQ/L	0.3995 - 0.4005	0.3997
Sample ID			
Tested 06/09/23			
Tester JENC			
Recorded 06/09/23			
QCSP5: STANDARDIZATION OF AQUEOUS ACIDS	YES METHOD		YES
Sample ID			
Tested 06/09/23			
Tester JENC			
Recorded 06/09/23			
N.I.S.T. Traceable to SRM 723 THAM	YES YES		YES
Sample ID			
Tested 06/09/23			
Tester JENC			
Recorded 06/09/23			
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 06/09/23			
Tester JENC			
Recorded 06/09/23			
SAMPLE COLLECTION & HANDLING PER LAB040	YES YES		YES
Sample ID			
Tested 06/09/23			
Tester JENC			
Recorded 06/09/23			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY MM/DD/YYYY		06/09/23
Sample ID			
Tested 06/09/23			
Tester JENC			
Recorded 06/09/23			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	21.8
Sample ID			
Tested 06/09/23			
Tester JENC			
Recorded 06/09/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		06/09/23
Sample ID			
Tested 06/09/23			
Tester JENC			
Recorded 06/09/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	21.8
DEG C	DEG C		
Sample ID			
Tested	06/09/23		
Tester	JENC		
Recorded	06/09/23		
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested	06/09/23		
Tester	JENC		
Recorded	06/09/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		GC,T-52A,BID-02253,PID-M91
Sample ID			
Tested	06/09/23		
Tester	JENC		
Recorded	06/09/23		
Lot #	3060914		
Made	06/09/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager