

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO # :

Packaged Product	3862		
	FLUORIDE STANDARD 500 PPM (	(1ML=0.5MG)	
Test	Target/UOM	Range	Result
PREPARED TO FORMU	ILATION ON YES		YES
FILE	YES		
Sample ID			
Tested	06/09/23		
Tester	CCOA		
Recorded	06/09/23		
APPEARANCE AND CO	LOR CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested	06/09/23		
Tester	CCOA		
Recorded	06/09/23		
INSTRUMENTS USED D	DURING INSTRUMENT		M91,VF2-B18
PREPARATION			
Sample ID			
Tested	06/09/23		
Tester	CCOA		
Recorded	06/09/23		
PPM FLUORIDE	500	490 - 510	496
	PPM		
Sample ID			
Tested	06/09/23		
Tester	CCOA		
Recorded	06/09/23		
EXPIRATION DATE	MM/DD/YYYY		12/30/23
	MM/DD/YYYY		
Sample ID			
Tested	06/09/23		
Tester	CCOA		
Recorded	06/09/23		
Lot # 30609	955		
Made 06/09	9/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA