

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8804 SODIUM THIOSULFATE 0.0394 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 06/05/23			
Tester JENC			
Recorded 06/05/23			
Actual Density	1.00277 GM/ML	1.00260 - 1.00293	1.00279
Sample ID			
Tested 06/05/23			
Tester JENC			
Recorded 06/05/23			
QCSP10:STANDARDIZATION OF SODIUM THIOSUL	YES QCSP10		YES
Sample ID			
Tested 06/05/23			
Tester JENC			
Recorded 06/05/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested 06/05/23			
Tester JENC			
Recorded 06/05/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M112,GC,,T269
Sample ID			
Tested 06/05/23			
Tester JENC			
Recorded 06/05/23			
Normality Reported to 5 decimal	0.03940 EQ/L	0.03932 - 0.03948	0.03944
Sample ID			
Tested 06/05/23			
Tester JENC			
Recorded 06/05/23			
PREPARED TO ASTM D 1510	YES YES		YES
Sample ID			
Tested 06/05/23			
Tester JENC			
Recorded 06/05/23			
SHELF LIFE OF PRODUCT	6 MONTHS TIME		6 MONTHS
Sample ID			
Tested 06/05/23			
Tester JENC			
Recorded 06/05/23			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 06/05/23 at 9:27 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

EXPIRATION DATE

MM/DD/YYYY

12/30/23

MM/DD/YYYY

Sample ID

Tested 06/05/23

Tester JENC

Recorded 06/05/23

Lot # 3053125

Made 06/05/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager