

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 1500			
BUFFER REFERENCE pH 7.0/YELLOW			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON</b>	YES		YES
<b>FILE</b>	YES		
Sample ID			
Tested 06/06/23			
Tester BPHI			
Recorded 06/06/23			
<b>PH OF PRODUCT/SOLUTION</b>	7.000 PH	6.990 - 7.010	6.996
Sample ID			
Tested 06/06/23			
Tester BPHI			
Recorded 06/06/23			
<b>+ PH Measurement Uncertainty</b>	0.014 PH	= 0.014	0.014
Sample ID			
Tested 06/06/23			
Tester BPHI			
Recorded 06/06/23			
<b>QCSP 20 PH CALIBRATION</b>	YES		YES
<b>METHODS</b>	METHOD		
Sample ID			
Tested 06/06/23			
Tester BPHI			
Recorded 06/06/23			
<b>Actual Temp</b>	25.0 DEG C	24.9 - 25.1	25.0
Sample ID			
Tested 06/06/23			
Tester BPHI			
Recorded 06/06/23			
<b>pH Traceable to N.I.S.T. SRM'S</b>	YES YES		YES
Sample ID			
Tested 06/06/23			
Tester BPHI			
Recorded 06/06/23			
<b>PH NIST TRACEABLE TO SRM 185</b>	YES YES		YES
Sample ID			
Tested 06/06/23			
Tester BPHI			
Recorded 06/06/23			
<b>NIST TRACEABLE TO SRM186I AND 186II</b>	YES YES		YES
Sample ID			
Tested 06/06/23			
Tester BPHI			
Recorded 06/06/23			

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>PH NIST TRACEABLE TO SRM 187</b>	YES YES	YES
Sample ID		
Tested	06/06/23	
Tester	BPHI	
Recorded	06/06/23	
<b>APPEARANCE AND COLOR</b>	CLEAR YELLOW/ODORLESS APPEARANCE	PASS
Sample ID		
Tested	06/06/23	
Tester	BPHI	
Recorded	06/06/23	
<b>EXPIRATION DATE</b>	MM/DD/YY MM/DD/YYYY	06/30/25
Sample ID		
Tested	06/06/23	
Tester	BPHI	
Recorded	06/06/23	
<b>SHELF LIFE OF PRODUCT</b>	2 YEARS TIME	2 YEARS
Sample ID		
Tested	06/06/23	
Tester	BPHI	
Recorded	06/06/23	
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT	M91,TIX
Sample ID		
Tested	06/06/23	
Tester	BPHI	
Recorded	06/06/23	
<b>Lot #</b>	30601105	
<b>Made</b>	06/06/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

  
**Quality Manager**