

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 6919			
CONDUCTIVITY STANDARD 100 uMHOS			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 06/13/23			
Tester BPHI			
Recorded 06/13/23			
CONDUCTIVITY @25 C	100	99 - 101	100
	US/CM		
Sample ID			
Tested 06/13/23			
Tester BPHI			
Recorded 06/13/23			
QCSP30:CONDUCTIVITY	YES		YES
STD&MEASUREMENT	QCSP30		
Sample ID			
Tested 06/13/23			
Tester BPHI			
Recorded 06/13/23			
TRACEABLE TO N.I.S.T. SRM 3191	YES		YES
	YES		
Sample ID			
Tested 06/13/23			
Tester BPHI			
Recorded 06/13/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 06/13/23			
Tester BPHI			
Recorded 06/13/23			
INSTRUMENTS USED DURING	INSTRUMENT		M92,T229
PREPARATION			
Sample ID			
Tested 06/13/23			
Tester BPHI			
Recorded 06/13/23			
EXPIRATION DATE	MM/DD/YY		06/30/24
	MM/DD/YYYY		
Sample ID			
Tested 06/13/23			
Tester BPHI			
Recorded 06/13/23			
SHELF LIFE OF PRODUCT	1 YEAR		1 YEAR
	TIME		
Sample ID			
Tested 06/13/23			
Tester BPHI			
Recorded 06/13/23			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 06/13/23 at 5:31 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Lot #	3061320
Made	06/13/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager