



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 7241			
POTASSIUM HYDROXIDE 0.1 NORMAL IN IPA (ASTM D664)			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 06/13/23			
Tester JENC			
Recorded 06/13/23			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 06/13/23			
Tester JENC			
Recorded 06/13/23			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		06/13/23
Sample ID			
Tested 06/13/23			
Tester JENC			
Recorded 06/13/23			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0	10.0 - 35.0	22.5
Sample ID			
Tested 06/13/23			
Tester JENC			
Recorded 06/13/23			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		06/13/23
Sample ID			
Tested 06/13/23			
Tester JENC			
Recorded 06/13/23			
<b>AMBIENT TEST TEMPERATURE DEGC</b>	20.0	15.0 - 30.0	22.5
Sample ID			
Tested 06/13/23			
Tester JENC			
Recorded 06/13/23			
<b>Actual Normality</b>	0.1000	0.0995 - 0.1005	0.1000
	EQ/L		
Sample ID			
Tested 06/13/23			
Tester JENC			
Recorded 06/13/23			
<b>QCSP7: STANDARDIZATION OF AQUEOUS BASES</b>	YES		YES
	METHOD		
Sample ID			
Tested 06/13/23			
Tester JENC			
Recorded 06/13/23			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>N.I.S.T. Traceable to SRM 84 (KHP)</b>	YES		YES
	YES		
Sample ID			
Tested	06/13/23		
Tester	JENC		
Recorded	06/13/23		
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE TO		PASS
	LIGHT HAZY		
	APPEARANCE		
Sample ID			
Tested	06/13/23		
Tester	JENC		
Recorded	06/13/23		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		T-58,M100,BID-02253,PIDM91
Sample ID			
Tested	06/13/23		
Tester	JENC		
Recorded	06/13/23		
<b>Actual Density</b>	0.79079	0.79022 - 0.79136	0.79095
	GM/ML		
Sample ID			
Tested	06/13/23		
Tester	JENC		
Recorded	06/13/23		
<b>SHELF LIFE OF PRODUCT</b>	SHELF LIFE		1 YEAR
	TIME		
Sample ID			
Tested	06/13/23		
Tester	JENC		
Recorded	06/13/23		
<b>Lot #</b>	3060946		
<b>Made</b>	06/13/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**