



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 1

Date: 06/15/23 at 5:35 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: SPX376F TOC STANDARD 375PPM (KHP)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 06/15/23			
Tester AROS			
Recorded 06/15/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested 06/15/23			
Tester AROS			
Recorded 06/15/23			
PRESERVED WITH HYDROCHLORIC ACID	YES YES		YES
Sample ID			
Tested 06/15/23			
Tester AROS			
Recorded 06/15/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M108,VF3-J9
Sample ID			
Tested 06/15/23			
Tester AROS			
Recorded 06/15/23			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		12/30/23
Sample ID			
Tested 06/15/23			
Tester AROS			
Recorded 06/15/23			
PPM TOTAL ORGANIC CARBON	375 MG/L (PPM)	369 - 381	371
Sample ID			
Tested 06/15/23			
Tester AROS			
Recorded 06/15/23			
Lot #	3061363		
Made	06/15/23		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 06/15/23 at 5:35 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Quality Manager