



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: HAC460F			
Calibration Soln 0.5 PPM CaCO3 for EZ1016-Total Hardness			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 06/20/23			
Tester AROS			
Recorded 06/20/23			
CALCULATED PPM CaCO3	0.50	0.47 - 0.53	0.50
	PPM		
Sample ID			
Tested 06/20/23			
Tester AROS			
Recorded 06/20/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	APPEARANCE		
Sample ID			
Tested 06/20/23			
Tester AROS			
Recorded 06/20/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P124
Sample ID			
Tested 06/20/23			
Tester AROS			
Recorded 06/20/23			
EXPIRATION DATE	MM/DD/YYYY		09/30/23
	MM/DD/YYYY		
Sample ID			
Tested 06/20/23			
Tester AROS			
Recorded 06/20/23			
Lot #	3061981		
Made	06/20/23		

Richard [Signature]

Quality Manager