



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 1526			
BUFFER REFERENCE pH 10.0/CLEAR			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 06/20/23			
Tester BPHI			
Recorded 06/20/23			
Actual pH of product	10.000 PH	9.990 - 10.010	9.996
Sample ID			
Tested 06/20/23			
Tester BPHI			
Recorded 06/20/23			
QCSP 20 PH CALIBRATION METHODS	YES METHOD		YES
Sample ID			
Tested 06/20/23			
Tester BPHI			
Recorded 06/20/23			
Actual Temp	25.0 DEG C	24.9 - 25.1	25.0
Sample ID			
Tested 06/20/23			
Tester BPHI			
Recorded 06/20/23			
N.I.S.T. Traceable to SRM 185	YES YES		YES
Sample ID			
Tested 06/20/23			
Tester BPHI			
Recorded 06/20/23			
NIST TRACEABLE TO SRM186I AND 186II	YES YES		YES
Sample ID			
Tested 06/20/23			
Tester BPHI			
Recorded 06/20/23			
N.I.S.T. TRACEABLE TO SRM 187	YES YES		YES
Sample ID			
Tested 06/20/23			
Tester BPHI			
Recorded 06/20/23			
TRACEABLE TO NIST SRM 191I AND SRM 191II	YES YES		YES
Sample ID			
Tested 06/20/23			
Tester BPHI			
Recorded 06/20/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested	06/20/23	
Tester	BPHI	
Recorded	06/20/23	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	T-IX,M87,M106
Sample ID		
Tested	06/20/23	
Tester	BPHI	
Recorded	06/20/23	
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	06/30/25
Sample ID		
Tested	06/20/23	
Tester	BPHI	
Recorded	06/20/23	
SHELF LIFE OF PRODUCT	SHELF LIFE TIME	2 YEARS
Sample ID		
Tested	06/20/23	
Tester	BPHI	
Recorded	06/20/23	
Lot #	3061511	
Made	06/20/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.


Quality Manager