



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8576			
SODIUM CARBONATE 1.0 NORMAL (0.5 Molar)			
Test	Target/UOM	Range	Result
Actual Normality	1.0000 EQ/L	0.9995 - 1.0005	0.9999
Sample ID			
Tested	06/19/23		
Tester	MHUS		
Recorded	06/19/23		
STANDARDIZED USING QCSP36	YES PASS		YES
Sample ID			
Tested	06/19/23		
Tester	MHUS		
Recorded	06/19/23		
N.I.S.T. Traceable to SRM 84	Yes YES		YES
Sample ID			
Tested	06/19/23		
Tester	MHUS		
Recorded	06/19/23		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	06/19/23		
Tester	MHUS		
Recorded	06/19/23		
Actual Density	1.05097 GM/ML	1.05048 - 1.05246	1.05132
Sample ID			
Tested	06/19/23		
Tester	MHUS		
Recorded	06/19/23		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	06/19/23		
Tester	MHUS		
Recorded	06/19/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M112,VF5-L16
Sample ID			
Tested	06/19/23		
Tester	MHUS		
Recorded	06/19/23		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		06/30/24
Sample ID			
Tested	06/19/23		
Tester	MHUS		
Recorded	06/19/23		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 06/19/23 at 5:18 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Lot #	3061660
Made	06/19/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager