



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: OXY203F			
LACHAT CALIBRATION STANDARD LEVEL 3			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 06/28/23			
Tester AROS			
Recorded 06/28/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested 06/28/23			
Tester AROS			
Recorded 06/28/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P-122,M100,VF2-A7,M108
Sample ID			
Tested 06/28/23			
Tester AROS			
Recorded 06/28/23			
PPM KIO3-NAOCL	40.0 PPM	39.9 - 40.1	40.0
Sample ID			
Tested 06/28/23			
Tester AROS			
Recorded 06/28/23			
CALCULATED GPL SODIUM HYDROXIDE	150.00 GPL	149.95 - 150.05	150.00
Sample ID			
Tested 06/28/23			
Tester AROS			
Recorded 06/28/23			
CALCULATED GPL SODIUM CHLORIDE	130.00 GPL	129.98 - 130.02	130.00
Sample ID			
Tested 06/28/23			
Tester AROS			
Recorded 06/28/23			
CALCULATED GPL SODIUM CHLORATE	1.0000 GPL	0.9995 - 1.0005	0.9999
Sample ID			
Tested 06/28/23			
Tester AROS			
Recorded 06/28/23			
SHELF LIFE OF PRODUCT	1 YEAR TIME		1 YEAR
Sample ID			
Tested 06/28/23			
Tester AROS			
Recorded 06/28/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

EXPIRATION DATE	MM/DD/YYYY		06/30/24
	MM/DD/YYYY		
Sample ID			
Tested	06/28/23		
Tester	AROS		
Recorded	06/28/23		
Actual Density	1.21720	1.21512 - 1.21928	1.21807
	GM/ML		
Sample ID			
Tested	06/28/23		
Tester	AROS		
Recorded	06/28/23		
Lot #	3062728		
Made	06/28/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager