

Page: 1

Date: 06/20/23 at 2:27 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	CY157F			
· ·	CYANIDE	STANDARD 100 PPM		
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	06/20/23			
Tester	JENC			
	06/20/23			
Actual PPM		100.00	98.00 - 102.00	99.69
		PPM		
Sample ID				
· ·	06/20/23			
	JENC			
	06/20/23			
QCSP47:STANDARDIZA		YES		YES
VARIOUS CYANIDE		QCSP47		. =0
Sample ID		Q001 17		
•	06/20/23			
	JENC			
	06/20/23			
APPEARANCE AND COL		CLEAR LIQUID		PASS
APPEARANCE AND COL	.On	APPEARANCE		FASS
0		AFFEARANCE		
Sample ID	00/00/00			
	06/20/23			
	JENC			
	06/20/23	INICTELINATION		D444 BID00050 TID44400
INSTRUMENTS USED DU	JRING	INSTRUMENT		P111,BID02253,TID41106
PREPARATION				
Sample ID				
	06/20/23			
	JENC			
	06/20/23			
SHELF LIFE OF PRODUC	CT	6 WEEKS		6 WEEKS
		TIME		
Sample ID				
Tested	06/20/23			
Tester	JENC			
	06/20/23			
EXPIRATION DATE		MM/DD/YYYY		08/01/23
		MM/DD/YYYY		
Sample ID				
	06/20/23			
	JENC			
	06/20/23			
Lot # 306192				
Made 06/20,	/ 23			



Page: 2

Date: 06/20/23 at 2:27 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager