



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

Test	Target/UOM	Range	Result
<b>Packaged Product:</b> SPE509F			
	TOTAL SUSPENDED SOLIDS STANDARD 1000 PPM		
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID	YES		
Tested 06/30/23			
Tester MMAT			
Recorded 06/30/23			
<b>APPEARANCE AND COLOR</b>	TURBID LIQUID APPEARANCE		PASS
Sample ID			
Tested 06/30/23			
Tester MMAT			
Recorded 06/30/23			
<b>CALCULATED MG/L (PPM) TSS</b>	1000.0 MG/L	995.0 - 1005.0	1000.0
Sample ID			
Tested 06/30/23			
Tester MMAT			
Recorded 06/30/23			
<b>ACT GMS CELITE</b>	1.0000 GM	0.9990 - 1.0010	1.0000
Sample ID			
Tested 06/30/23			
Tester MMAT			
Recorded 06/30/23			
<b>QCSP11:STANDARDIZATION OF BROMIDE-BROMAT</b>	YES QCSP11		YES
Sample ID			
Tested 06/30/23			
Tester MMAT			
Recorded 06/30/23			
<b>MLS OF WATER</b>	1000 ML	995 - 1005	1000
Sample ID			
Tested 06/30/23			
Tester MMAT			
Recorded 06/30/23			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		VF1-Z33
Sample ID			
Tested 06/30/23			
Tester MMAT			
Recorded 06/30/23			
<b>WEIGHT NIST TRACEABLE</b>	PASS YES		PASS
Sample ID			
Tested 06/30/23			
Tester MMAT			
Recorded 06/30/23			



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)


**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>EXPIRATION DATE</b>	MM/DD/YY MM/DD/YYYY	07/30/24
Sample ID		
Tested	06/30/23	
Tester	MMAT	
Recorded	06/30/23	
<b>SHELF LIFE OF PRODUCT</b>	<b>SHELF LIFE</b>	<b>1 YEAR</b>
	<b>TIME</b>	
Sample ID		
Tested	06/30/23	
Tester	MMAT	
Recorded	06/30/23	
<b>Lot #</b>	3063008	
<b>Made</b>	06/30/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



**Quality Manager**