

Page: 1

Date: 06/29/23 at 11:09 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc	t: DC562F			
		T 91 HYDROLYZABLE CHLORI	DE REACTANT	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES	-	YES
FILE		YES		
Sample ID)			
Tested	06/29/23			
Tester	MMAT			
Recorded	06/29/23			
PPM LITHIUM CHLORIDE W/V)		5.50	5.30 - 5.70	5.63
	•	PPM		
Sample ID)			
Tested	06/29/23			
Tester	MMAT			
Recorded				
Actual Density	00/23/20	0.90023	0.89791 - 0.90254	0.89997
Actual Delisity		GM/ML	0.03731 - 0.30234	0.09391
Camanda ID		GIVI/IVIL		
Sample ID				
Tested	06/29/23			
Tester	MMAT			
Recorded		0.545.05440544045		B. 60
APPEARANCE AND COLOR		CLEAR ORANGE LIQUID		PASS
		APPEARANCE		
Sample ID)			
Tested	06/29/23			
Tester	MMAT			
Recorded	06/29/23			
INSTRUMENTS USED	DURING	INSTRUMENT		M108,VF4-N31,M100,GC
PREPARATION				
Sample ID)			
Tested	06/29/23			
Tester	MMAT			
Recorded	06/29/23			
EXPIRATION DATE		MM/DD/YYYY		07/30/24
		MM/DD/YYYY		
Sample ID)			
Tested	06/29/23			
Tester	MMAT			
Recorded	06/29/23			
Lot # 3062				
Made 06/2				
wade 00/2	J / 4J			



Page: 2

Date: 06/29/23 at 11:09 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager