

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 9764-10L			
WATER, DEIONIZED REAGENT GRADE ASTM TYPE III (10LC)			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON</b>	YES		YES
<b>FILE</b>	YES		
Sample ID			
Tested 06/30/23			
Tester TBUR			
Recorded 06/30/23			
<b>PREPARED TO ASTM D 1193</b>	PASS		PASS
	PASS		
Sample ID			
Tested 06/30/23			
Tester TBUR			
Recorded 06/30/23			
<b>CONDUCTIVITY</b>	<= 0.25 UMHOS		<=0.25 UMHOS
	US/CM		
Sample ID			
Tested 06/30/23			
Tester TBUR			
Recorded 06/30/23			
<b>RESISTIVITY 18 MEG-OHM AT TIME</b>	>= 4.0 MEGOHM		>= 4.0 MEGOHM
<b>OF PACKAG</b>	PASS		
Sample ID			
Tested 06/30/23			
Tester TBUR			
Recorded 06/30/23			
<b>TOC &lt; 200 ppb</b>	<= 200 PPB		<= 200 PPB
	< 200 PPB		
Sample ID			
Tested 06/30/23			
Tester TBUR			
Recorded 06/30/23			
<b>Sodium (Na) &lt;=10 ppb</b>	<= 10 PPB		<=10 PPB
	<= 10 PPB		
Sample ID			
Tested 06/30/23			
Tester TBUR			
Recorded 06/30/23			
<b>Chloride (Cl) &lt;= 10 ppb</b>	<= 10 PPB		<= 10 PPB
	<= 10 PPB		
Sample ID			
Tested 06/30/23			
Tester TBUR			
Recorded 06/30/23			
<b>Silica &lt;= 500 ppb</b>	<= 500 PPB		<= 500 PPB
	<= 500 PPB		
Sample ID			
Tested 06/30/23			
Tester TBUR			
Recorded 06/30/23			


**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>MEETS ASTM D 1193 TYPE III SPECIFICATION</b>	PASS	PASS
Sample ID		
Tested 06/30/23		
Tester TBUR		
Recorded 06/30/23		
<b>SHELF LIFE OF PRODUCT</b>	1 YEAR	1 YEAR
	TIME	
Sample ID		
Tested 06/30/23		
Tester TBUR		
Recorded 06/30/23		
<b>EXPIRATION DATE</b>	MM/DD/YY	05/30/24
	MM/DD/YYYY	
Sample ID		
Tested 06/30/23		
Tester TBUR		
Recorded 06/30/23		
<b>Lot #</b> 3052267		
<b>Made</b> 05/24/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



**Quality Manager**